

Chroma Meter CS-200 Communication Specifications



<p>Chroma Meter CS-200</p> <p>Communication Specifications</p>
--

First edition:	January 26, 2005
Second edition:	March 12, 2006; for instrument ROM ver. 1.10
Third edition:	April 12, 2006; Corrections to second edition
Fourth edition:	August 4, 2008; Addition of notes regarding use with a PC equipped with an OHCI-standard USB host controller. (MES, MDR)
Fifth edition	August 19, 2013 Changes company name

KONICA MINOLTA, INC.

Chroma Meter CS-200 Communication Specifications

Notes concerning software and associated documentation:

- When Chroma Meter CS-200 is used connected with a PC or other devices, USB driver for CS-200 is needed.
- USB driver for CS-200 and/or accompanying manuals (hereinafter referred to as "SOFTWARE") is copyrighted by Konica Minolta. Alteration of this software without the express permission of Konica Minolta is strictly prohibited.
- Unauthorized copying of this software, in whole or in part, is strictly prohibited.
- This software can be used by customers who purchased Chroma Meter CS-200 and agree with the terms of the End User License Agreement.
- The commands and other information described in the Communication specifications cannot be used for purposes related to products for commercial sales without the express permission of Konica Minolta, Inc.
- This document is designed for users who are familiar with the fundamentals of PC communications.
- Reproduction of this document, its contents, or any portion thereof without the permission of Konica Minolta, Inc. is strictly prohibited.
- This document is subject to change without notice.
- Konica Minolta, Inc. and/or its distributors are not liable to damages, problems, or loss, which may be incurred as a result of using this software.
- Every effort has been taken to ensure the accuracy of the information in this document, if there are any questions or if any errors are found, please feel free to contact the nearest Konica Minolta representative.
- Company names and product names that appear within this document are trademarks or registered trademarks of their respective companies.

Chroma Meter CS-200 Communication Specifications

Table of Contents

BLR (Read current display backlight setting)	6
BLS (Turn display backlight on/off).....	7
CAD (Delete all calibration channel data)	8
CDR (Read CH data)	9
CDS (Set backed-up CH data on the unit)	11
CHD (Delete specific calibration channel data).....	13
CHR (Read current calibration channel settings)	14
CHS (Set calibration channel)	15
CIS (Set calibration channel ID).....	16
CSR (Read color space setting)	17
CSS (Set color space)	18
DMR (Read display mode setting)	19
DMS (Set display mode)	20
DPR (Read data protection mode)	21
DPS (Activate data protection)	22
DTR (Read current date/time setting).....	23
DTS (Set current date/time setting)	24
ETR (Calculate and apply calibration coefficient).....	25
FSR (Read operation status of the finder shutter)	26
FSS (Set current operation status of the finder shutter).....	27
IDR (Read ID data of the unit)	28
LNR (Read lens type and angle of measurement)	29
LNS (Set lens)	30
MAD (Delete all measurement data from the memory).....	31
MDD (Delete specific measurement data from the memory)	32
MDR (Read measurement data setting)	33
MEM (Save measurement data)	35
MES (Start/stop measurement)	36
MMR (Read current save mode setting).....	37
MMS (Set save mode)	38
MSR (Read measurement mode setting).....	39
MSS (Set measurement mode).....	40
OBR (Read current Observer setting)	41
OBS (Set Observer setting)	42
RMT (Enable/disable remote mode)	43

Chroma Meter CS-200 Communication Specifications

SCR (Read current synchronization mode setting)	44
SCS (Set synchronization mode).....	45
SDR (Retrieve measurement data from the memory).....	46
SFR (Read number of significant figures for chromaticity)	48
SFS (Set number of significant figures for chromaticity)	49
SIS (Assign ID for measurement data in the memory)	50
SMR (Read sleep mode setting)	51
SMS (Turn on/off sleep mode)	52
SPR (Read current measurement speed setting).....	53
SPS (Set measurement speed)	54
STR (Read current status of the unit).....	55
SWE (Enable/disable trigger mode in the unit).....	56
TAD (Delete all target data).....	57
TDD (Delete specific target data)	58
TDR (Retrieve target data from the memory)	59
TDS (Set target data)	62
UCS (Enable/disable calibration mode)	63
UWT (Perform user calibration)	64
ZRC (Perform zero-calibration).....	65

Chroma Meter CS-200 Communication Specifications

1. Overview

- All commands from a PC to the CS-200 and the responses from the CS-200 are in fixed-length ASCII format. (Please refer to 4. Character Code Table.) All command sequences need to be followed by the delimiter code (CR+LF).

Outbound Data (from PC to CS-200): Maximum 64 characters

Inbound Data (from CS-200 to PC): Fixed at 250 characters

- A comma is used to separate a command, Error Check Code, and each associated parameter.
- When the Error Check Code is "ER**" (where ** is a number), parameters will not be returned.
- All commands and responses are case-sensitive. A space is processed as a character.

Chroma Meter CS-200 Communication Specifications

2. Communication Commands

BLR (Read current display backlight setting)

Command Name	BLR: <u>B</u> ack <u>l</u> ight <u>R</u> ead
Function	Read current display backlight setting
<p>Operating Procedure:</p> <p>[PC] [Unit]</p> <p>"BLR"+ delimiter----></p> <p style="text-align: right;"><-----" <i>Error Check Code, Mode Setting</i> "+ delimiter</p>	
<p>Description:</p> <p>This function allows users to read the current display backlight setting.</p>	
<p>Format:</p> <p>- <i>Mode Setting</i> ""</p>	
<p>Parameter Values:</p> <p>- <i>Mode Setting</i> "0": OFF "1": ON</p>	

Chroma Meter CS-200 Communication Specifications

CAD (Delete all calibration channel data)

Command Name	CAD: <u>C</u> hannel Data <u>A</u> ll <u>D</u> elete
Function	Delete all calibration channel data
Operating Procedure:	
[PC]	[Unit]
"CAD"+ delimiter ---->	<-----" <i>Error Check Code</i> "+ delimiter
Description:	
This function allows users to delete the stored data for all calibration channels (ID, calibration coefficients, & target color).	
Format:	
Parameter Values:	

Chroma Meter CS-200 Communication Specifications

CDR (Read CH data)

Command Name	CDR: <u>C</u> hannel <u>D</u> ata <u>R</u> ead
Function	Read CH data (to back up data)
<p>Operating Procedure:</p> <p>[PC] [Unit]</p> <p>"CDR, <i>Calibration CH</i>" + [delimiter]----></p> <p><-----"Error Check Code, CH data, Calibration coefficients" + [delimiter]</p> <p>"CNT" + [delimiter]----></p> <p><-----"Error Check Code, Target values, Calibration values" + [delimiter]</p>	
<p>Description:</p> <p>This function allows users to make a backup of CH data on a PC. The data can then be copied or used to restore CH data in the unit by using the CDS command.</p> <p>CDR and CNT responses constitute one set of channel data. Therefore, the CNT command should follow immediately after receiving the response to the CDR command. If another command intervenes in the sequence, the command to read the calibration channel data will be canceled, and the process must be started again from the beginning.</p> <p>Note:</p> <p>Calibration CH00 is the calibration channel for the Konica Minolta preset calibration standard, and contains the compensation coefficients for the Konica Minolta calibration standard. These data cannot be changed. As a result, although data for CH00 can be read from the instrument using this command, the data for CH00 cannot be set on the instrument using the CDS command.</p>	
<p>Format:</p> <ul style="list-style-type: none"> - <i>Calibration CH</i> "***" - <i>CH data</i> "***, *, *, *****" (<i>Calibration CH, Mode, Observer, ID</i>) - <i>Calibration coefficients</i> "*****, *****,, *****" (16 characters × 9 sets) (<i>a11, a12, a13, a21, a22, a23, a31, a32, a33</i>) - <i>Target values</i> "*****, *****, *****" (<i>X, Y, Z</i>) - <i>Calibration values</i> "*****, *****, *****,, *****" (11 characters × 12 sets) (<i>X_W, Y_W, Z_W, X_R, Y_R, Z_R, X_G, Y_G, Z_G, X_B, Y_B, Z_B</i>) 	
<p>Parameter Values:</p> <ul style="list-style-type: none"> - <i>Calibration CH</i> " 0" to "20" 	

Chroma Meter CS-200 Communication Specifications

[CH data]	
- <i>Calibration CH</i>	"_0" to "20"
- <i>Mode</i>	"0": Konica Minolta preset standard "1": Single-point calibration "2": Matrix calibration
- <i>Observer</i>	"0": 2° "1": 10° "9": Same as Observer configuration set on unit
- <i>ID data</i>	9 single-byte characters
- <i>Calibration coefficients</i>	"0000000000000000" to "FFFFFFFFFFFFFFFF" (IEEE double precision)
- <i>Target values</i>	
- <i>Calibration values</i>	

Chroma Meter CS-200 Communication Specifications

CDS (Set backed-up CH data on the unit)

Command Name	CDS: <u>C</u> hannel <u>D</u> ata <u>S</u> et
Function	Set backed-up CH data on the unit
Operating Procedure:	
[PC]	[Unit]
"CDS, <i>CH Data</i> " + delimiter ----	<-----"Error Check Code" + delimiter
"CNT, <i>Calibration coefficients 1</i> " + delimiter ----	<-----"Error Check Code" + delimiter
"CNT, <i>Calibration coefficients 2</i> " + delimiter ----	<-----"Error Check Code" + delimiter
"CNT, <i>Calibration coefficients 3</i> " + delimiter ----	<-----"Error Check Code" + delimiter
"CNT, <i>Target values</i> " + delimiter ----	<-----"Error Check Code" + delimiter
"CNT, <i>Calibration values 1</i> " + delimiter ----	<-----"Error Check Code" + delimiter
"CNT, <i>Calibration values 2</i> " + delimiter ----	<-----"Error Check Code" + delimiter
"CNT, <i>Calibration values 3</i> " + delimiter ----	<-----"Error Check Code" + delimiter
"CNT, <i>Calibration values 4</i> " + delimiter ----	<-----"Error Check Code" + delimiter
Description:	
<p>This function allows users to set previously backed up calibration channel data on the unit. The data should be sent via the CNT command immediately after executing the CDS command. If another command intervenes in the sequence, the command to set up the calibration channel data will be canceled.</p>	
Caution:	
<p>Calibration CH00 is the calibration channel for the Konica Minolta preset calibration standard, and contains the compensation coefficients for the Konica Minolta calibration standard. These data cannot be changed. As a result, although data for CH00 can be read from the instrument using the CDR command, the data for CH00 cannot be set on the instrument using the CDS command.</p>	
<p>In addition, calibration data are crucial for computing measured values. The data</p>	

Chroma Meter CS-200 Communication Specifications

retrieved using the CDR command should be set again on the unit without alterations.

Format:

- *CH data* **"** , * , * , *****"**
 (*Calibration CH, Mode, Observer, ID*)
- *Calibration coefficients 1* **"***** , ***** , *****"**
 (*a₁₁, a₁₂, a₁₃*)
- *Calibration coefficients 2* **"***** , ***** , *****"**
 (*a₂₁, a₂₂, a₂₃*)
- *Calibration coefficients 3* **"***** , ***** , *****"**
 (*a₃₁, a₃₂, a₃₃*)
- *Target values* **"***** , ***** , *****"**
 (*X, Y, Z*)
- *Calibration values 1* **"***** , ***** , *****"**
 (*X_W, Y_W, Z_W*)
- *Calibration values 2* **"***** , ***** , *****"**
 (*X_R, Y_R, Z_R*)
- *Calibration values 3* **"***** , ***** , *****"**
 (*X_G, Y_G, Z_G*)
- *Calibration values 4* **"***** , ***** , *****"**
 (*X_B, Y_B, Z_B*)

Parameter Values:

- *Calibration CH* **"_1" to "20"**

Chroma Meter CS-200 Communication Specifications

CIS (Set calibration channel ID)

Command Name	CIS: Calibration <u>C</u> hannel I <u>I</u> dentification <u>S</u> et
Function	Set calibration channel ID
Operating Procedure: <div style="display: flex; justify-content: space-between; margin-top: 10px;"> [PC] [Unit] </div> <p>"CIS, <i>Calibration CH</i>, <i>ID</i>" + delimiter----> <div style="text-align: right; margin-top: 10px;"><-----" <i>Error Check Code</i>" + delimiter</div> </p>	
Description: This function allows users to set a calibration channel ID consisting of 9 single-byte characters of fixed length. Refer to "4. Character Code Table" for the characters that can be used for entry.	
Format: - <i>Calibration CH</i> "***" - <i>ID</i> "*****"	
Parameter Values: - <i>Calibration CH</i> "_1" to "20" - <i>ID</i> 9 single-byte characters or less	

Chroma Meter CS-200 Communication Specifications

CSR (Read color space setting)

Command Name	CSR: Display <u>C</u> olor <u>S</u> pace Mode <u>R</u> ead		
Function	Read color space setting		
Operating Procedure:			
[PC]	[Unit]		
"CSR"+ delimiter ---->			
<-----"Error Check Code, Color space"+ delimiter			
Description:			
This function allows users to read the color space setting on the display.			
Format:			
- <i>Color space</i> "*"			
Parameter Values:			
- <i>Color space</i> "0": L _v x _y "1": L _v u'v' "2": L _v TΔu _v			
"3": XYZ "4": Dominant wavelength			

Chroma Meter CS-200 Communication Specifications

CSS (Set color space)

Command Name	CSS: Display <u>C</u> olor <u>S</u> pace Mode <u>S</u> et		
Function	Set color space		
Operating Procedure:			
[PC]		[Unit]	
"CSS, <i>Color space</i> " + delimiter ----> <div style="text-align: right;"><-----"Error Check Code" + delimiter</div>			
Description:			
This function allows users to set the color space.			
Caution:			
When 10° observer is selected, "LvTΔuv" cannot be selected.			
If the display mode is set to Ratio (%X%Y%Z) and the color space is changed to a space other than XYZ, the display mode will change to Color difference. In such case, even if the color space is changed back to XYZ, the display mode will remain as Color difference (ΔXΔYΔZ) and will not automatically change back to Ratio.			
Format:			
- <i>Color space</i>	"*"		
Parameter Values:			
- <i>Color space</i>	"0": Lvxy	"1": Lvu'v'	"2": LvTΔuv
	"3": XYZ	"4": Dominant wavelength	

Chroma Meter CS-200 Communication Specifications

DTS (Set current date/time setting)

Command Name	DTS: <u>D</u> ate/ <u>T</u> ime <u>S</u> et
Function	Set current date/time setting
Operating Procedure: <div style="display: flex; justify-content: space-between; margin-top: 10px;"> [PC] [Unit] </div> <p>"DTS, <i>Year/Month/Day, Hour/Minute/Second</i>"</p> <p style="text-align: center;">+ delimiter----></p> <p style="text-align: right;"><-----"Error Check Code"+ delimiter</p>	
Description: This function allows users to set the current date/time setting. Numerals between "00" (2000) and "99" (2099) can be entered for the year.	
Format: - <i>Year/Month/Day</i> "*****" - <i>Hour/Minute/Second</i> "*****"	
Parameter Values: - <i>Year/Month/Day</i> "00_1_1" to "991231" (YY/MM/DD) - <i>Hour/Minute/Second</i> "_0_0_0" to "235959"	

Chroma Meter CS-200 Communication Specifications

ETR (Calculate and apply calibration coefficient)

Command Name	ETR: <u>E</u> <u>n</u> <u>t</u> <u>e</u> <u>r</u>
Function	Calculate and apply calibration coefficient
Operating Procedure: <div style="display: flex; justify-content: space-between; align-items: center;"> [PC] [Unit] </div> <div style="display: flex; justify-content: space-between; align-items: center;"> "ETR"+ delimiter----> <-----" <i>Error Check Code</i>" + delimiter </div>	
Description: This function allows for the calculation and application of the calibration coefficient. The coefficient takes effect when the unit is set to the calibration mode. The computation is made based on the values stored at this point and the result is registered in the unit. The command will fail if the calibration mode is released before executing this command. User calibration should be executed in the following order: UCS-UWT-ETR-UCS (release). The channel subject to calibration is read with the CHR ("CHR,99") command (and changed with the CHS command).	
Format:	
Parameter Values:	

Chroma Meter CS-200 Communication Specifications

IDR (Read ID data of the unit)

Command Name	IDR: <u>I</u> dentification <u>R</u> ead
Function	Read ID data of the unit
Operating Procedure: <div style="display: flex; justify-content: space-between; margin-top: 10px;"> [PC] [Unit] </div> <div style="margin-top: 10px;"> "IDR"+ delimiter----> <div style="text-align: right; margin-top: 10px;"> <-----"Error Check Code, ID data"+ delimiter </div> </div>	
Description: This function allows users to read the ID data of the unit (i.e. product type, ROM version number and product number).	
Format: <div style="margin-top: 10px;"> - <i>ID data</i> "*****, **, *****" <div style="text-align: center; margin-top: 10px;"> (Product type, ROM version number, Product number) </div> </div>	
Parameter Values: <div style="margin-top: 10px;"> - <i>Product type</i> Fixed to "1892-100" - <i>ROM version number</i> "***": 3 characters (ver*.**) </div> <div style="margin-top: 10px;"> - <i>Product number</i> "*****": 7 characters </div>	

Chroma Meter CS-200 Communication Specifications

LNS (Set lens)

Command Name	LNS: Measurement Lens Set
Function	Set lens
Operating Procedure:	
[PC]	[Unit] "LNS, Lens"+ [delimiter]----> <div style="text-align: right;"><-----"Error Check Code"+ [delimiter]</div>
Description: This function allows users to set a lens for measurement. Coefficients for the lenses, filters and others need to be registered in a PC software program before selecting any of the options from User Lens 1 to User Lens 3.	
Format: - Lens "*"	
Parameter Values: - Lens "0": Standard lens "1": Close-up lens 1 (No. 107) "2": Close-up lens 2 (No. 122) "3": User Lens 1 (When lens coefficient is set up by PC software.) "4": User Lens 2 (When lens coefficient is set up by PC software.) "5": User Lens 3 (When lens coefficient is set up by PC software.)	

Chroma Meter CS-200 Communication Specifications

MDD (Delete specific measurement data from the memory)

Command Name	MDD: <u>M</u> easurement <u>D</u> ata <u>D</u> ele
Function	Delete specific measurement data from the memory
Operating Procedure: [PC] [Unit] "MDD, <i>Data number</i> " + delimiter ---> <div style="text-align: right;"><-----"Error Check Code" + delimiter</div>	
Description: This function allows users to delete specific measurement data that has been stored in the memory.	
Format: - <i>Data number</i> "****"	
Parameter Values: - <i>Data number</i> "__0" to "100"	

Chroma Meter CS-200 Communication Specifications

MDR (Read measurement data setting)

Command Name	MDR: <u>M</u> easurement <u>D</u> ata <u>R</u> ead
Function	Read measurement data setting
Operating Procedure:	
[PC]	[Unit] "MDR, <i>Color space</i> " + [delimiter]----> <-----" <i>Error Check Code, Measurement value data, Measured values</i> " + [delimiter]
Description: This function allows users to read the measurement data setting. The value for Lv is expressed according to the unit of luminance selected for the device. (only when Observer is set to 2°) Caution: If the value for Lv, X, Y, or Z exceeds display range, error code of "OK12" or "OK13" is returned together with "-9999999999" as the corresponding measurement value parameter. If the value for T or Δuv exceeds display range, the corresponding parameter will be blank. When performing measurements using a computer equipped with an OHCI-standard USB host controller, attention should be paid to the following when controlling the CS-200. When using the MDR command to obtain data after taking a measurement, the MDR command should be sent 0.5s before the end of the measurement. (For example, a wait time of [measurement time-0.5]s should be inserted between the MES and MDR commands.) Then, if the response to the MDR command is “ER02”, the program should wait 0.3s before resending the MDR command. For a sample program, refer to “Chroma Meter CS-200 DLL Reference”.	
Format:	
- <i>Color space</i>	"*"
- <i>Measurement value data</i>	"*, *, *, **, *, *****, *, **" (Lens, Angle of measurement, Measurement speed, Duration of measurement, Synchronization, Frequency, Observer, Calibration CH)
- <i>Measured values</i>	"***** , ***** , *****"
Parameter Values:	
- <i>Color space</i>	"0": LvxY "1": Lvu'v' "2": LvTΔuv "3": XYZ "4": Dominant wavelength

Chroma Meter CS-200 Communication Specifications

[Measurement Value Data]

- Lens

"0": Standard lens

"1": Close-up lens 1 (No. 107)

"2": Close-up lens 2 (No. 122)

"3": User Lens 1 (When lens coefficient is set up by PC software.)

"4": User Lens 2 (When lens coefficient is set up by PC software.)

"5": User Lens 3 (When lens coefficient is set up by PC software.)

- Angle of measurement

"0": 0.1°

"1": 0.2°

"2": 1°

- Measurement speed

"0": LTD. AUTO

"1": S-FAST

"2": FAST

"3": SLOW

"4": S-SLOW

"5": MANU

"6": AUTO

- Duration of measurement (seconds) "_1" to "60"

- Synchronization "0": No-Sync "1": Sync

- Frequency "____0" or "_4000" to "20000" (Frequency × 100)

- Observer "0": 2° "1": 10°

- Calibration channel "_0" to "20"

[Measured values]

Responses are expressed according to the color space specified in the command parameter.

(Ex) Lvxv

"____80.003, ____0.3127, ____0.3293"

(Lv)

(x)

(y)

(Ex) Lvu'v'

"____80.003, ____0.3333, ____0.3333"

(Lv)

(u')

(v')

(Ex) LvTΔuv (When Δuv is a positive value, the first character is left blank.)

"____80.003, ____6500, ____-0.0050"

(Lv)

(T)

(Δuv)

(Ex) XYZ

"____55.442, ____80.003, ____9.001"

(X)

(Y)

(Z)

(Ex) Dominant wavelength

"____80.003, ____550.4, _____"

(Lv)

(Dominant wavelength)

(Blank)

Chroma Meter CS-200 Communication Specifications

MEM (Save measurement data)

Command Name	MEM: <u>MEM</u> ory
Function	Save measurement data
Operating Procedure:	
[PC]	[Unit]
"MEM, <i>Data number</i> " +	delimiter ---->
	<-----" <i>Error Check Code</i> " + delimiter
Description:	
This function allows users to save measurement data.	
Format:	
- <i>Data number</i>	"***"
Parameter Values:	
- <i>Data number</i>	"__0" to "100"

Chroma Meter CS-200 Communication Specifications

MES (Start/stop measurement)

Command Name	MES: <u>M</u> ea <u>s</u> ure
Function	Start/stop measurement
<p>Operating Procedure:</p> <p>[PC] [Unit]</p> <p>"MES, <i>Mode setting</i>" + delimiter----></p> <p style="text-align: right;"><-----"Error Check Codes, Duration of measurement" + delimiter</p>	
<p>Description:</p> <p>This function allows users to start/stop measurement. The MDR command should be used to retrieve the measurement data.</p> <p>Caution:</p> <p>If the value for Lv, X, Y, or Z exceeds display range, error code of "OK12" or "OK13" is returned together with "-9999999999" as the corresponding measurement value parameter. If the value for T or Δuv exceeds display range, the corresponding parameter will be blank. When performing measurements using a computer equipped with an OHCI-standard USB host controller, attention should be paid to the following when controlling the CS-200. When using the MDR command to obtain data after taking a measurement, the MDR command should be sent 0.5s before the end of the measurement. (For example, a wait time of [measurement time-0.5]s should be inserted between the MES and MDR commands.) Then, if the response to the MDR command is “ER02”, the program should wait 0.3s before resending the MDR command.</p> <p>For a sample program, refer to “Chroma Meter CS-200 DLL Reference”.</p>	
<p>Format:</p> <p>- <i>Mode setting</i> "*)"</p> <p>- <i>Duration of measurement</i> "***"</p>	
<p>Parameter Values:</p> <p>- <i>Mode setting</i> "1": Start measurement "0": Stop measurement</p> <p>- <i>Duration of measurement</i> "_0" or "_1" to "60"</p>	

Chroma Meter CS-200 Communication Specifications

MMR (Read current save mode setting)

Command Name	MMR:
Function	Read current save mode setting
Operating Procedure:	
[PC]	[Unit]
"MMR"+ delimiter ----->	<-----" <i>Error Check Code, Save mode</i> "+ delimiter
Description:	
This function allows users to read the current save mode setting when the unit is used as a freestanding device. The measurement data is saved when the MEMORY key is depressed.	
Format:	
- <i>Save mode</i>	"*"
Parameter Values:	
-- <i>Save mode</i>	
"0": AUTO_NUM	"1": MAN_NUM "2": AUTOSAVE

Chroma Meter CS-200 Communication Specifications

MMS (Set save mode)

Command Name	MMS:
Function	Set save mode
Operating Procedure: [PC] [Unit] "MMS, <i>Save mode</i> " + delimiter ----> <-----" <i>Error Check Code</i>" + delimiter	
Description: This function allows users to set the save mode when the unit is used as a freestanding device. The measurement data is saved when the MEMORY key is depressed.	
Format: - <i>Save mode</i> "*"	
Parameter Values: - <i>Save mode</i> "0": AUTO_NUM "1": MAN_NUM "2": AUTOSAVE	

Chroma Meter CS-200 Communication Specifications

MSR (Read measurement mode setting)

Command Name	MSR: <u>M</u> ea <u>s</u> Mode <u>R</u> ead
Function	Read measurement mode setting
Operating Procedure: <div style="display: flex; justify-content: space-between; align-items: flex-start;"> <div style="text-align: left;"> [PC] "MSR" + delimiter----> </div> <div style="text-align: right;"> [Unit] <-----"Error Check Code, Mode setting"+ delimiter </div> </div>	
Description: This function allows users to read the measurement mode setting.	
Format: - <i>Mode setting</i> "*"	
Parameter Values: - <i>Mode setting</i> "0": Single measurement "1": Continuous measurement	

Chroma Meter CS-200 Communication Specifications

MSS (Set measurement mode)

Command Name	MSS: <u>M</u> ea <u>s</u> Mode <u>S</u> et
Function	Set measurement mode
Operating Procedure: <div style="display: flex; justify-content: space-between; margin-top: 10px;"> [PC] [Unit] </div> <div style="margin-top: 10px;"> "MSS, <i>Mode setting</i>" + delimiter----> <div style="text-align: right; margin-right: 100px;"> <-----" <i>Error Check Code</i>" + delimiter </div> </div>	
Description: This function allows users to set up the measurement mode. The setting is only applicable to the unit when it is used as a freestanding device. The unit is only capable of taking single measurements when connected to a computer in remote mode.	
Format: - <i>Mode setting</i> "*"	
Parameter Values: - <i>Mode setting</i> "0": Single measurement "1": Continuous measurement	

Chroma Meter CS-200 Communication Specifications

OBR (Read current Observer setting)

Command Name	OBR: <u>O</u> bserver <u>R</u> ead
Function	Read current Observer setting
Operating Procedure:	
[PC]	[Unit]
"OBR"+ <u>delimiter</u> ----	<-----" <i>Error Check Code, Observer</i> "+ <u>delimiter</u>
Description:	
This function allows users to read the current Observer setting.	
Format:	
- <i>Observer</i>	""
Parameter Values:	
- <i>Observer</i>	"0": 2° "1": 10°

Chroma Meter CS-200 Communication Specifications

OBS (Set Observer setting)

Command Name	OBS: <u>O</u> bserver <u>S</u> et
Function	Set Observer setting
<p>Operating Procedure:</p> <p>[PC] [Unit]</p> <p>"OBS, <i>Observer</i>" + delimiter----></p> <p style="text-align: right;"><-----" <i>Error Check Code</i>" + delimiter</p>	
<p>Description:</p> <p>This function allows users to set the Observer setting.</p>	
<p>Format:</p> <p>- <i>Observer</i> "*" "</p>	
<p>Parameter Values:</p> <p>- <i>Observer</i> "0": 2° "1": 10°</p>	

Chroma Meter CS-200 Communication Specifications

RMT (Enable/disable remote mode)

Command Name	RMT: <u>R</u> emote <u>M</u> ode Set
Function	Enable/disable remote mode
Operating Procedure: <div style="display: flex; justify-content: space-between; align-items: flex-start;"> <div style="width: 45%;"> [PC] "RMT, <i>Mode setting</i>" + delimiter----> </div> <div style="width: 45%;"> [Unit] <-----"Error Check Code" + delimiter </div> </div>	
Description: This function allows users to enable/disable remote mode. When the remote mode is ON, various commands can be given via communication but not via the unit keys. When the remote mode is OFF, the unit is unable to communicate with other devices ("ER16" appears), but it can be controlled using the unit keys.	
Format: - <i>Mode setting</i> "*"	
Parameter Values: - <i>Mode setting</i> "0": Remote mode is OFF. (Communication function OFF/Unit key control ON) "1": Remote mode is ON. (Communication function ON/Unit key control OFF)	

Chroma Meter CS-200 Communication Specifications

SCR (Read current synchronization mode setting)

Command Name	SCR: <u>S</u> ynchronization Mode <u>R</u> ead
Function	Read current synchronization mode setting
Operating Procedure:	
[PC]	[Unit]
"SCR"+ delimiter ----	<-----" <i>Error Check Code, Synchronization, Frequency</i> "+ delimiter
Description:	
This function allows users to read the current synchronization mode setting.	
Format:	
- <i>Synchronization</i>	"*"
- <i>Frequency</i>	"*****"
Parameter Values:	
- <i>Synchronization</i>	"0": No-Sync "1": Sync
- <i>Frequency (Hz)</i>	"____0": No setting "_4000" to "20000": Frequency (40.0Hz to 200.0Hz) × 100

Chroma Meter CS-200 Communication Specifications

SCS (Set synchronization mode)

Command Name	SCS: <u>S</u> ynchronization Mode <u>S</u> et
Function	Set synchronization mode
Operating Procedure: <div style="display: flex; justify-content: space-between; margin-top: 10px;"> [PC] [Unit] </div> <p>"SCS, <i>Synchronization, Frequency</i>" + delimiter----> <-----" <i>Error Check Code</i>" + delimiter</p>	
Description: This function allows users to set the synchronization mode. If No-Sync is selected, a measurement is completed within the timeframe apt for the speed of measurement. In the Sync mode, the duration of the measurement is determined based upon the frequency that has been selected. Caution: Frequency is automatically set to "0" if No-sync is selected. "0" cannot be set as the frequency if Sync is selected.	
Format: <div style="margin-top: 5px;"> - <i>Synchronization</i> "*" </div> <div style="margin-top: 5px;"> - <i>Frequency</i> "*****" </div>	
Parameter Values: <div style="margin-top: 5px;"> - <i>Synchronization</i> "0": No-Sync "1": Sync </div> <div style="margin-top: 5px;"> - <i>Frequency (Hz)</i> "____0": No setting "4000" to "20000": Frequency (40.0Hz to 200.0Hz) × 100 </div>	

Chroma Meter CS-200 Communication Specifications

SDR (Retrieve measurement data from the memory)

Command Name	SDR: <u>S</u> tored <u>D</u> ata <u>R</u> equest
Function	Retrieve measurement data from the memory
Operating Procedure:	
[PC]	[Unit]
"SDR, <i>Data number, Color space</i> " + [delimiter]----> <-----" <i>Error Check Code, Measurement value data, Measured values</i> " + [delimiter]	
Description:	
This function allows users to download specific measurement data that have been stored in the unit to a PC.	
The value for Lv is expressed according to the unit of luminance selected for the device (only when the Observer is set to 2°).	
Caution:	
If the value for Lv, X, Y, or Z exceeds display range, error code of "OK12" or "OK13" is returned together with "-9999999999" as the corresponding measurement value parameter.	
If the value for T or Δuv exceeds display range, the corresponding measurement parameter will be blank.	
Format:	
- <i>Data number</i>	"****"
- <i>Color space</i>	"**"
- <i>Measurement value data</i>	"*, *, *, **, *, *****, *, **, ***** , ***** , *****"
(Lens, Angle of measurement, Measurement speed, Duration of measurement, Synchronization, Frequency, Observer, Calibration CH, Measurement Data ID, Date of measurement, Time of measurement)	
- <i>Measured values</i>	"***** , ***** , *****"
Parameter Values:	
- <i>Data number</i>	"__0" to "100"
- <i>Color space</i>	"0": L _{vxy} "1": L _{vu'v'} "2": L _{vTΔuv} "3": XYZ "4": Dominant wavelength
[<i>Measurement value data</i>]	
- Lens	
"0": Standard lens	
"1": Close-up lens 1 (No. 107)	
"2": Close-up lens 2 (No. 122)	
"3": User Lens 1 (When lens coefficient is set up by PC software.)	

Chroma Meter CS-200 Communication Specifications

"4": User Lens 2 (When lens coefficient is set up by PC software.)

"5": User Lens 3 (When lens coefficient is set up by PC software.)

- Angle of measurement

"0": 0.1°

"1": 0.2°

"2": 1°

- Measurement speed

"0": LTD. AUTO

"1": S-FAST

"2": FAST

"3": SLOW

"4": S-SLOW

"5": MANU

"6": AUTO

- Duration of measurement (seconds) "_1" to "60"

- Synchronization "0": No-Sync "1": Sync

- Frequency "___0" No setting
 "_4000" to "20000" (Frequency × 100)

- Observer "0": 2° "1": 10°

- Calibration CH "_0" to "20"

- Measurement data ID (Please refer to 4. Character Code Table.)

- Date of measurement "00_1_1" to "991231" (YY/MM/DD)

- Time of measurement "_0_0_0" to "235959"

[Measured values]

Responses are expressed according to the color space specified in the command parameter.

(Ex) Lvx_y

"___80.003, ___0.3127, ___0.3293"
 (Lv) (x) (y)

(Ex) Lvu'v'

"___80.003, ___0.3333, ___0.3333"
 (Lv) (u') (v')

(Ex) LvT Δ uv (When Δ uv is a positive value, the first character is left blank.)

"___80.003, ___6500, ___-0.0050"
 (Lv) (T) (Δ uv)

(Ex) XYZ

"___55.442, ___80.003, ___9.001"
 (X) (Y) (Z)

(Ex) Dominant wavelength

"___80.003, ___550.4, _____"
 (Lv) (Dominant wavelength) (Blank)

Chroma Meter CS-200 Communication Specifications

SFS (Set number of significant figures for chromaticity)

Command Name	SFS: <u>S</u> ignificant <u>F</u> igure <u>S</u> et	
Function	Set number of significant figures for chromaticity	
Operating Procedure:	<p>[PC] [Unit]</p> <p>"SFS, <i>Number of significant figures</i>" + delimiter----></p> <p style="text-align: right;"><-----" <i>Error Check Code</i>" + delimiter</p>	
Description:	This function allows users to set the number of significant figures for chromaticity.	
Format:	<p>- <i>Number of significant figures</i> "*" "</p>	
Parameter Values:	<p>- <i>Number of significant figures</i> "0": 3 characters "1": 4 characters</p>	

Chroma Meter CS-200 Communication Specifications

SIS (Assign ID for measurement data in the memory)

Command Name	SIS: <u>S</u> tored Data <u>I</u> dentification <u>S</u> et
Function	Assign ID for measurement data in the memory
Operating Procedure: <div style="display: flex; justify-content: space-between; margin-top: 10px;"> [PC] [Unit] </div> <p>"SIS, <i>Data number</i>, <i>ID</i>" + delimiter----></p> <p style="text-align: right;"><-----" <i>Error Check Code</i>" + delimiter</p>	
Description: This function allows users to assign an ID for the measurement data stored in the memory. The ID should be entered with 9 single-byte fixed-length characters. Refer to "4. Character Code Table" for the characters that can be used for entry.	
Format: - <i>Data number</i> "****" - <i>ID</i> "*****"	
Parameter Values: - <i>Data number</i> "__0" to "100" - <i>ID</i> 9 single-byte characters	

Chroma Meter CS-200 Communication Specifications

SMR (Read sleep mode setting)

Command Name	SMR: <u>S</u> leep <u>M</u> ode <u>R</u> ead
Function	Read sleep mode setting
Operating Procedure: <div style="display: flex; justify-content: space-between; align-items: flex-start;"> <div style="text-align: left;"> [PC] "SMR + delimiter----> </div> <div style="text-align: right;"> [Unit] <-----"Error Check Code, Mode setting"+ delimiter </div> </div>	
Description: This function allows users to read the sleep mode setting.	
Format: - <i>Mode setting</i> "*"	
Parameter Values: - <i>Mode setting</i> "0": OFF "1": ON	

Chroma Meter CS-200 Communication Specifications

SMS (Turn on/off sleep mode)

Command Name	SMS: <u>S</u> leep <u>M</u> ode <u>S</u> et
Function	Turn on/off sleep mode
Operating Procedure: <div style="display: flex; justify-content: space-between; align-items: flex-start;"> <div style="text-align: left;">[PC]</div> <div style="text-align: right;">[Unit]</div> </div> <div style="display: flex; justify-content: space-between; align-items: flex-start; margin-top: 10px;"> <div style="text-align: left;"> "SMS, <i>Mode setting</i>" + delimiter----> </div> <div style="text-align: right;"> <-----" <i>Error Check Code</i>" + delimiter </div> </div>	
Description: This function allows users to turn the sleep mode on/off.	
Format: - <i>Mode setting</i> "*"	
Parameter Values: - <i>Mode setting</i> "0": OFF "1": ON	

Chroma Meter CS-200 Communication Specifications

SPR (Read current measurement speed setting)

Command Name	SPR: Measurement <u>S</u> peed <u>R</u> ead			
Function	Read current measurement speed setting			
Operating Procedure:				
[PC]	[Unit]			
"SPR"+ [delimiter]---->	<-----"Error Check Code, Measurement speed, Duration of measurement"+ [delimiter]			
Description:				
This function allows users to read the current measurement speed setting.				
Format:				
- <i>Measurement speed</i>	**"			
- <i>Duration of measurement</i>	***"			
Parameter Values:				
- <i>Measurement speed</i>				
"0": LTD. AUTO	"1": S-FAST	"2": FAST	"3": SLOW	
"4": S-SLOW	"5": MANU	"6": AUTO		
- <i>Duration of measurement</i>				
"_1" to "60": Duration of measurement				
" 0": No setting				

Chroma Meter CS-200 Communication Specifications

SPS (Set measurement speed)

Command Name	SPS: Measurement <u>S</u> peed <u>S</u> et
Function	Set measurement speed
Operating Procedure: [PC] [Unit] "SPS, <i>Measurement speed, Duration of measurement</i> " + delimiter ----> <-----" <i>Error Check Code</i> " + delimiter	
Description: This function allows users to set the measurement speed. Caution: Duration of measurement can be set up only when MANU is selected.	
Format: - <i>Measurement speed</i> "*" - <i>Duration of measurement</i> "***"	
Parameter Values: - <i>Measurement speed</i> "0": LTD. AUTO "1": S-FAST "2": FAST "3": SLOW "4": S-SLOW "5": MANU "6": AUTO - <i>Duration of measurement</i> "_1" to "60": Duration of measurement "_0": No setting	

Chroma Meter CS-200 Communication Specifications

STR (Read current status of the unit)

Command Name	STR: <u>S</u> tatus <u>R</u> equest
Function	Read current status of the unit
Operating Procedure:	
[PC]	[Unit]
"STR"+ [delimiter]---->	<-----"Error Check Code, Status"+ [delimiter]
Description:	This function allows users to read the current status of the unit.
Format:	- <i>Status</i> " <u>*</u> , <u>**</u> , <u>***</u> " (Battery status, Number of measurement data in memory, Maximum number of measurement data)
Parameter Values:	- <i>Battery status</i> "0": Full "1": Close to exhausted - <i>Number of measurement data in memory</i> "__0" to "101" - <i>Maximum number of measurement data</i> "101"

Chroma Meter CS-200 Communication Specifications

SWE (Enable/disable trigger mode in the unit)

Command Name	SWE: Measurement <u>S</u> witch <u>E</u> nable
Function	Enable/disable trigger mode in the unit
Operating Procedure: <div style="display: flex; justify-content: space-between;"> [PC] [Unit] </div> <div style="display: flex; justify-content: space-between;"> "SWE, <i>Mode setting</i>" + delimiter---- > </div> <div style="display: flex; justify-content: space-between;"> <-----"Error Check Code" + delimiter </div>	
Description: This function allows users to enable/disable the trigger mode in the unit. When the trigger mode is ON, measurements can be taken only using the unit keys, overriding the remote mode. When the trigger mode is turned OFF, the unit returns to remote mode. Caution: When the trigger mode is changed from OFF to ON, hold data stored in the instrument will be deleted. Data measured while the trigger mode is ON can be read from the instrument using the MDR command. When the data is read using the MDR command, the hold data will be deleted after reading.	
Format: - <i>Mode setting</i> "*"	
Parameter Values: - <i>Mode setting</i> "0": Trigger mode is OFF. (Remote mode) "1": Trigger mode is ON. (Remote mode/Unit key control is ON)	

Chroma Meter CS-200 Communication Specifications

TAD (Delete all target data)

Command Name	TAD: <u>T</u> arget Data <u>A</u> ll <u>D</u> elete
Function	Delete all target data
Operating Procedure: <div style="display: flex; justify-content: space-around; margin-top: 10px;"> [PC] [Unit] </div> <div style="margin-top: 20px;"> "TAD"+ delimiter----> <div style="text-align: right; margin-top: 10px;"> <-----"Error Check Code"+ delimiter </div> </div>	
Description: This function allows users to delete all target data that have been stored in the memory.	
Format:	
Parameter Values:	

Chroma Meter CS-200 Communication Specifications

TDD (Delete specific target data)

Command Name	TDD: <u>T</u> arget <u>D</u> ata <u>D</u> elete
Function	Delete specific target data
<p>Operating Procedure:</p> <p>[PC] [Unit]</p> <p>"TDD, <i>Data number</i>" + <u>delimiter</u>----></p> <p><-----" <i>Error Check Code</i>" + <u>delimiter</u></p>	
<p>Description:</p> <p>This function allows users to delete specific target data that have been stored in the memory.</p>	
<p>Format:</p> <p>- <i>Data number</i> "***"</p>	
<p>Parameter Values:</p> <p>- <i>Data number</i> "_1" to "20"</p>	

Chroma Meter CS-200 Communication Specifications

TDR (Retrieve target data from the memory)

Command Name	TDR: <u>T</u> arget <u>D</u> ata <u>R</u> equest
Function	Retrieve target data from the memory
Operating Procedure:	
[PC]	[Unit]
"TDR, <i>Data number, Color space</i> " + [delimiter]----> <-----"Error Check Code, Target value data, Target value" + [delimiter]	
Description:	
This function allows users to retrieve specific target data that has been stored in the unit and download the data to a PC.	
The value for Lv is expressed according to the unit of luminance selected for the device. (only when Observer is set to 2°)	
When the target value is manually entered, the data type is expressed as "1": Entered data, in which case, the entries for lens, angle of measurement, measurement speed, duration of measurement, synchronization and frequency are left blank.	
Caution:	
If value for Lv, X, Y, or Z exceeds display range, error code of "OK12" or "OK13" is returned together with "-9999999999" as the corresponding target value parameter.	
If the value for T or Δuv exceeds the display range, the corresponding target value parameter will be blank.	
Format:	
- Data number	***"
- Color space	***"
- Target value data	", *, *, *, **, *, *****, *, **, *****", *****, *****" (Data type, Lens, Angle of measurement, Measurement speed, Duration of measurement, Synchronization, Frequency, Observer, Calibration CH, Target value ID, Date of measurement, Time of measurement)
- Target value	"*****", *****", *****" (Data type, Lens, Angle of measurement, Measurement speed, Duration of measurement, Synchronization, Frequency, Observer, Calibration CH, Target value ID, Date of measurement, Time of measurement)
Parameter Values:	
- Data number	"_1" to "20"
- Color space	"0": LvxY "1": Lvu'v' "2": LvTΔuv "3": XYZ "4": Dominant wavelength

Chroma Meter CS-200 Communication Specifications

[Target value data]

- Data type

"0": Measured data

"1": Entered data

- Lens

"0": Standard lens

"1": Close-up lens 1 (No. 107)

"2": Close-up lens 2 (No. 122)

"3": User Lens 1 (When lens coefficient is set up by PC software.)

"4": User Lens 2 (When lens coefficient is set up by PC software.)

"5": User Lens 3 (When lens coefficient is set up by PC software.)

- Angle of measurement

"_" or "0": 0.1°

"1": 0.2°

"2": 1°

- Measurement speed

"_" or "0": LTD. AUTO

"1": S-FAST

"2": FAST

"3": SLOW

"4": S-SLOW

"5": MANU

"6": AUTO

- Duration of measurement (seconds)

"_" or "_1" to "60"

- Synchronization

"_" or "0": No-Sync

"1": Sync

- Frequency

"_" or "_0": No setting

"_4000" to "20000": Frequency × 100

- Observer

"0": 2°

"1": 10°

- Calibration CH

"_0" to "20"

- Measurement data ID (Please refer to 4. Character Code Table.)

- Date of measurement "_00_1_1" to "991231" (YY/MM/DD)

- Time of measurement "_0_0_0" to "235959"

[Target value]

Responses are expressed according to the color space specified in the command parameter.

(Ex) L_vx_y

"_80.003, _0.3127, _0.3293"

(L_v)

(x)

(y)

(Ex) L_{vu}'v'

"_80.003, _0.3333, _0.3333"

(L_v)

(u')

(v')

(Ex) L_vTΔ_{uv} (When Δ_{uv} is a positive value, the first character is left blank.)

"_80.003, _6500, _-0.0050"

(L_v)

(T)

(Δ_{uv})

Chroma Meter CS-200 Communication Specifications

(Ex) XYZ

"____.55.442, ____80.003, ____9.001"

(X)

(Y)

(Z)

(Ex) Dominant wavelength

"____80.003, ____550.4, ____"

(L_v)

(Dominant wavelength)

(Blank)

Chroma Meter CS-200 Communication Specifications

TDS (Set target data)

Command Name	TDS: <u>T</u> arget <u>D</u> ata <u>S</u> et
Function	Set target data
Operating Procedure: <div style="display: flex; justify-content: space-between;"> [PC] [Unit] </div> <p>"TDS, <i>Data number</i>, <i>Color space</i>, <i>Target data</i>" + delimiter----> <-----" <i>Error Check Code</i>" + delimiter</p>	
Description: This function allows users to set target data for the data number specified. This data is used to calculate color difference.	
Format: <div style="margin-left: 20px;"> - <i>Data number</i> "***" - <i>Color space</i> "*" " - <i>Target value</i> "***** , ***** , *****" </div>	
Parameter Values: <div style="margin-left: 20px;"> - <i>Data number</i> "_1" to "20" - <i>Color space</i> "0": L_vx_y "1": L_vu'v' "3": XYZ - <i>Target value</i> (Enter the parameter values specific to the color space) </div> <div style="margin-left: 20px;"> (Ex) L_vx_y <div style="display: flex; justify-content: space-around; width: 100%;"> "____80.003, ____0.3127, ____0.3293" </div> <div style="display: flex; justify-content: space-around; width: 100%;"> (L_v) (x) (y) </div> </div> <div style="margin-left: 20px;"> (Ex) L_vu'v' <div style="display: flex; justify-content: space-around; width: 100%;"> "____80.003, ____0.3333, ____0.3333" </div> <div style="display: flex; justify-content: space-around; width: 100%;"> (L_v) (u') (v') </div> </div> <div style="margin-left: 20px;"> (Ex) XYZ <div style="display: flex; justify-content: space-around; width: 100%;"> "____55.442, ____80.003, ____9.001" </div> <div style="display: flex; justify-content: space-around; width: 100%;"> (X) (Y) (Z) </div> </div>	

* Parameter values of color space (target or calibration value) should be set with the following range.

$$0 < x < 1 \text{ and}$$

$$0 < y < 1 \text{ and}$$

$$0 < x+y \leq 1 \text{ and}$$

$$0 < X \leq 99999000000 \text{ and}$$

$$0 < Y \leq 99999000000 \text{ and}$$

$$0 \leq Z \leq 99999000000$$

Chroma Meter CS-200 Communication Specifications

UCS (Enable/disable calibration mode)

Command Name	UCS: <u>U</u> ser <u>C</u> alibration <u>S</u> etting
Function	Enable/disable calibration mode
Operating Procedure: <div style="display: flex; justify-content: space-between;"> [PC] [Unit] </div> <div style="display: flex; justify-content: space-between;"> "UCS, <i>Mode</i>" + delimiter----> <-----" <i>Error Check Code</i>" + delimiter </div>	
Description: This function allows users to enable/disable user calibration mode. This command is unavailable for CH00. User calibration should be executed in the following order: UCS-UWT-ETR-UCS (release). The channel subject to calibration is read with the CHR ("CHR,99") command (and changed with the CHS command).	
Format: - <i>Mode setting</i> "*"	
Parameter Values: - <i>Mode setting</i> "0": OFF "1": ON	

Chroma Meter CS-200 Communication Specifications

UWT (Perform user calibration)

Command Name	UWT: <u>U</u> ser Calibration <u>W</u> hite
Function	Perform single-point user calibration
Operating Procedure: <div style="display: flex; justify-content: space-between;"> [PC] [Unit] </div> <p style="text-align: center;"> "UWT, <i>Data number</i>, <i>Color space</i>, <i>Calibration value</i>" + delimiter----> <-----" <i>Error Check Code</i>" + delimiter </p>	
Description: This function allows users to perform single-point user calibration. User calibration should be executed in the following order: UCS-UWT-ETR-UCS (release). The channel subject to calibration is read with the CHR ("CHR,99") command (and changed with the CHS command).	
Format: - <i>Data number</i> "****" - <i>Color space</i> "*" - <i>Calibration value</i> "***** , ***** , *****"	
Parameter Values: - <i>Data number</i> "0" to "100" or "999": the value on hold - <i>Color space</i> "0": L _v x _y "1": L _v u'v' "3": XYZ - <i>Calibration value</i> (Enter the parameter values specific to the color space) (Ex) L _v x _y <div style="display: flex; justify-content: space-around; margin-top: 10px;"> " ____ 80.003, ____ 0.3127, ____ 0.3293" </div> <div style="display: flex; justify-content: space-around; margin-top: 10px;"> (L_v) (x) (y) </div> (Ex) L _v u'v' <div style="display: flex; justify-content: space-around; margin-top: 10px;"> " ____ 80.003, ____ 0.3333, ____ 0.3333" </div> <div style="display: flex; justify-content: space-around; margin-top: 10px;"> (L_v) (u') (v') </div> (Ex) XYZ <div style="display: flex; justify-content: space-around; margin-top: 10px;"> " ____ 55.442, ____ 80.003, ____ 9.001" </div> <div style="display: flex; justify-content: space-around; margin-top: 10px;"> (X) (Y) (Z) </div>	

* Parameter values of color space (target or calibration value) should be set with the following range.

$$0 < x < 1 \text{ and}$$

$$0 < y < 1 \text{ and}$$

$$0 < x+y \leq 1 \text{ and}$$

$$0 < X \leq 99999000000 \text{ and}$$

$$0 < Y \leq 99999000000 \text{ and}$$

$$0 \leq Z \leq 99999000000$$

Chroma Meter CS-200 Communication Specifications

ZRC (Perform zero-calibration)

Command Name	ZRC: <u>Z</u> ero <u>C</u> alibration
Function	Perform zero-calibration
Operating Procedure:	
[PC]	[Unit]
"ZRC"+ delimiter ----	<-----" <i>Error Check Code</i> "+ delimiter
Description: This function allows users to perform zero-calibration.	
Format:	
Parameter Values:	

Chroma Meter CS-200 Communication Specifications

3. Error Check Codes

Codes	Description
OK00	OK
OK03	Low battery voltage (The battery voltage is between 3.4V-3.6V)
OK12	Lv, X, Y, or Z exceeded display range
OK13	Low battery voltage (The battery voltage is between 3.4V-3.6V) and Lv, X, Y, or Z exceeded display range
ER01	Low battery voltage (The battery voltage is below 3.4V)
ER02	No command is accepted while taking measurements
ER03	Invalid entry for Lvxy or Lvu'y' during user calibration value or target value setup
ER05	Invalid entry for matrix calibration
ER06	Invalid matrix coefficient entry during matrix calibration (The elements on the main diagonal are below 0. The determinant is 0.)
ER07	CH00 does not accept user setup
ER08	Incorrect observer setting. Ex: Observer setting for the CH selected differs from observer setting of the instrument, the color space of "LvTΔuv" is selected when 10° observer was selected, etc.
ER09	Data protection is ON. MEM command to save the current data failed.
ER10	No command
ER11	Inbound data exceeds 64 characters
ER14	Incorrect parameter format
ER15	Incorrect parameter range or parameter error
ER16	Invalid operation of communication commands. For example: Instrument isn't in remote mode; user calibration procedure not completed correctly; sending CNT command with CDR or CDS isn't appropriate; etc.
ER20	No data
ER21	Low luminance
ER22	Beyond the range of measurement
ER23	Offset error (Shutter error)
ER27	Unstable range (due to excessive luminance variation)
ER30	Position of measuring angle selector isn't correct. The angle of measurement was changed during measurement.
ER31	FROM write error
ER34	Clock IC error
ER35	A/D conversion error

Chroma Meter CS-200 Communication Specifications

4. Character Code Table

The following characters can be used to assign IDs for measurement values and calibration channels.

Sp indicates a space code. Codes with blank entries are unavailable.

	20	30	40	50	60	70
0	sp	0	@	P	'	p
1	!	1	A	Q	a	q
2	"	2	B	R	b	r
3	#	3	C	S	c	s
4	\$	4	D	T	d	t
5	%	5	E	U	e	u
6	&	6	F	V	f	v
7	'	7	G	W	g	w
8	(8	H	X	h	x
9)	9	I	Y	i	y
A	*	:	J	Z	j	z
B	+	;	K	[k	{
C	,	<	L	\	l	
D	-	=	M]	m	}
E	.	>	N	^	n	
F	/	?	O	_	o	